



Certificate of Analysis



| Material | Lot | Batch | Expiration date* |
|---------------------------------------|-----|-------|------------------|
| Silica Gel 20/40 Specially Cleaned | | | |

* This expiration date is for sealed tubes. Tubes that are open-ended and made from this batch will have a shorter shelf life.

This batch of sorbent has met our QC requirements as outlined in SKC Procedure 4022.

Ryan P. Sojak

Ryan P. Sojak
Quality Assurance Manager

| Analytical Results | |
|--------------------|----------------|
| Anion | µg/400 mg |
| Fluoride | < LOD (3.0 µg) |
| Chloride | < LOD (3.0 µg) |
| Bromide | < LOD (3.0 µg) |
| Nitrate | < LOD (3.0 µg) |
| Phosphate | < LOD (3.0 µg) |
| Sulfate | < LOD (3.0 µg) |