



Certificate of Analysis

Material	Lot	Batch	Expiration date*
XAD-2	12885	MKCK4106	January, 2025

* This expiration date is for sealed tubes. Tubes that are open-ended and made from this batch will have a shorter shelf life.

Scott Marshall
QA Manager

Product Name:
Supelpak™ -2 - matrix styrene-divinylbenzene, 20-60 mesh, jar of 1000 g

Product Number: 21130-U
Batch Number: MKCK4106
Brand: SUPELCO
CAS Number: 9003-70-7
Quality Release Date: 25 NOV 2019

Test	Specification	Result
Appearance (Color)	White to Faint Yellow and Faint Beige and Faint Brown	White
Appearance (Form)	Powder or Solid or Beads	Beads
Blank Recovery	60.0 - 120.0 %	75.8 %
By GC		
Sample Recovery	60.0 - 120.0 %	102.2 %
By GC		
Sample Background	≤ 10.0	< 0.0
(µg/g) By GC		
Peak Area	Pass	Pass