



Certificate of Analysis

Material	Lot	Batch	Expiration date*
XAD-2	13124	MKCM0907	June, 2025

* This expiration date is for sealed tubes. Tubes that are open-ended and made from this batch will have a shorter shelf life.

Scott Marshall

Quality Assurance Manager

Product Name:

Supelpak™-2 - matrix styrene-divinylbenzene, 20-60 mesh, jar of 1000 g

Product Number: 21130-U

Batch Number: MKCM0907

Brand: SUPELCO

CAS Number: 9003-70-7

Quality Release Date: 22 APR 2020

Test	Specification	Result
Appearance (Color)	White to Faint Yellow and Faint Beige and Faint Brown	White
Appearance (Form)	Powder or Solid or Beads	Beads
Blank Recovery	60.0 - 120.0 %	83.0 %
By GC		
Sample Recovery	60.0 - 120.0 %	92.9 %
By GC		
Sample Background	≤ 10.0	3.4
(µg/g) By GC		
Peak Area	Pass	Pass
Resin Largest Single Peak Area < Average		
Standard Heptane Peak Area		