



## Certificate of Analysis

Material	Lot	Batch	Expiration date*
XAD-2	13125	MKCM0906	June, 2025

\* This expiration date is for sealed tubes. Tubes that are open-ended and made from this batch will have a shorter shelf life.

Scott Marshall

Quality Assurance Manager

Product Name:  
Supelpak™-2 - matrix styrene-divinylbenzene, 20-60 mesh, jar of 1000 g

Product Number: 21130-U  
Batch Number: MKCM0906  
Brand: SUPELCO  
CAS Number: 9003-70-7  
Quality Release Date: 02 APR 2020

Test	Specification	Result
Appearance (Color)	White to Faint Yellow and Faint Beige and Faint Brown	White
Appearance (Form)	Powder or Solid or Beads	Beads
Blank Recovery	60.0 - 120.0 %	94.9 %
By GC		
Sample Recovery	60.0 - 120.0 %	79.2 %
By GC		
Sample Background	≤ 10.0	1.1
(µg/g) By GC		
Peak Area	Pass	Pass
Resin Largest Single Peak Area < Average		
Standard Heptane Peak Area		