



863 Valley View Road, Eighty Four, PA 15330
724-941-9701
skcinc.com

Coated Filters
Cat. No. 225-9011
Operating Instructions

Method & Chemical: MDHS 25/3 (UK) for Organic Isocyanates; PI, TDI, HDI, MDI.

Filter & Coating: 25-mm glass fiber filter coated with 1-(2-methoxyphenyl)piperazine. Filters only, not loaded in cassettes

Prior to Sampling: Store at ≤ 39.2 F (4 C). Limited shelf-life; check expiration date on packaging.

Sample Stability: Store TDI at ≤ 39.2 F (4 C) for up to 90 days. No data is available for PI, HDI, and MDI.

For information on other available coated filters, go to www.skcinc.com.

How to Use SKC Coated Filters



Ensure that filters are at room temperature before sampling.

1. Set up the sampling train for flow rate verification with the representative filter cassette in line. For “closed-face” sampling, remove the plugs and connect the cassette to the sampling train. For “open-face” sampling, remove the outlet plug and the cassette inlet section and then connect the cassette to the sampling train. For details on setting up a sampling train, refer to the SKC Sample Setup Guide “Sampling Train – Filters” at www.skccinc.com/knowledgecenter.
2. Verify the pump flow rate using a flowmeter.
3. Replace the cassette used to verify flow rate with a fresh filter cassette for sample collection. Attach the cassette to a worker’s collar, as close to the breathing zone as possible, by using a filter cassette holder SKC Cat. No. 225-1. Sample for the specified time interval and record the time.
4. Remove the filter cassette at the end of the specified sampling period and replace both end plugs (and the inlet if necessary). Recheck the flow rate using the same cassette and flowmeter used in Step 2 to ensure that the flow rate has not changed by more than 5%.
5. Store sample at ≤ 39.2 F (4 C) for up to 90 days. Appropriately package samples, blanks, and all pertinent data to send to a laboratory for analysis. Expedite shipping of samples to laboratory. Friday shipments are not recommended.

SKC Limited Warranty and Return Policy

SKC products are subject to the SKC Limited Warranty and Return Policy, which provides SKC’s sole liability and the buyer’s exclusive remedy. To view the complete SKC Limited Warranty and Return Policy, go to skccinc.com/warranty.