



863 Valley View Road, Eighty Four, PA 15330  
724-941-9701  
skcinc.com

**Li-Ion Battery Pack for AirChek® TOUCH, Essential,  
Essential+, and Connect Sample Pumps  
Cat. No. P75718  
Operating Instructions**

**Notes and Cautions**

- Power off pump before removing battery to avoid loss of time, date, and other settings.
- Use only the SKC-approved battery pack (Cat. No. P75718) and charging cradle (Cat. No. 220-800 or 220-900) for pump. Use of an unapproved battery and/or charging cradle could damage the pump and will void any warranty.
- Tampering with the battery pack (opening, disassembling, short circuiting, crushing, or exposing the battery pack to fire or temperatures in excess of 212 F [100 C]) voids any warranty.
- Failure to follow warnings, notes, and cautions voids any warranty.
- WARNING: Substitution of components may impair intrinsic safety.  
AVERTISSEMENT: La substitution de composants peut compromettre la Sécurité Intrinsèque.
- CAUTION: The battery used in this device may present a risk of fire or explosion when heated above 212 F (100 C) or incinerated. Replace battery with SKC Unlisted Component Battery Pack model P75718 only. Use of another battery may present a risk of fire or explosion.
- WARNING: To prevent ignition of a hazardous atmosphere, batteries must only be changed [removed and replaced] in an area known to be non-hazardous.  
AVERTISSEMENT: Afin de prévenir l'inflammation d'atmosphères dangereuses, ne changer les batteries que dans des emplacements désignés non dangereux.
- CAUTION: Risk of Fire and Burns. Do Not Disassemble, heat above 212 F (100 C), or incinerate. Keep battery out of reach of children and in original package until ready to use. Dispose of used batteries promptly according to [all state and] local recycling or waste regulations.

## Replacing the Battery Pack



**Review Notes and Cautions on page 1 before proceeding.**



**Turn pump off before removing battery pack.**

1. Turn pump off by pressing and holding briefly the recessed power button on the side of the pump (screen turns off).
2. Remove the existing battery pack.
  - a. Use a 2.5-mm hex driver (Allen wrench) to loosen two screws on bottom of battery pack housing.
  - b. Pull battery pack housing away from pump case.
  - c. If replacing battery pack with a new battery pack Cat. No. P75718, dispose of the used battery promptly.



**Do not disassemble the battery pack. Do not dispose of in fire. Dispose of used batteries promptly according to all state and local recycling or waste regulations.**

3. Install a new battery pack or reinstall existing battery pack.
  - a. Align battery pack with bottom of pump case. **Note:** *The connector on top of battery pack should align with protruding power control board contacts on bottom of pump case.*
  - b. Press the two parts together until snug.
  - c. Use a 2.5-mm hex driver (Allen wrench) to tighten two screws on bottom of battery pack housing. Tighten screws in alternating fashion.
  - d. Charge the battery pack completely before use. If reinstalling existing battery pack, ensure that it is charged to at least 25% (battery status icon on Home screen shows two bars).

### **SKC Limited Warranty and Return Policy**

SKC products are subject to the SKC Limited Warranty and Return Policy, which provides SKC's sole liability and the buyer's exclusive remedy. To view the complete SKC Limited Warranty and Return Policy, go to [skcinc.com/warranty](http://skcinc.com/warranty).