

CERTIFICATE OF COMPLIANCE

Certificate Number 20141105-E62011
Report Reference E62011-20141031
Issue Date 2014-NOVEMBER-05

Issued to: SKC INC
863 VALLEYVIEW RD
EIGHTY FOUR PA 15330

**This is to certify that
representative samples of**

AIR-SAMPLING EQUIPMENT FOR USE IN HAZARDOUS
LOCATIONS

USL, CNL Intrinsically safe portable air sampling pump
AirChek Touch Model 220-5000TC for use in Class I, Div. 1,
Groups A, B, C, and D; Class II, Div. 1, Groups E, F, and G;
and Class III, Div. 1 hazardous locations when used with
SKC Battery Pack Model P75718.

Have been investigated by UL in accordance with the
Standard(s) indicated on this Certificate.

Standard(s) for Safety:


Standard No. UL 913 - Intrinsically Safe Apparatus and
Associated Apparatus for Use in Class I, II, and III, Division
1, Hazardous (Classified) Locations
Standard No. CAN/CSA C22.2 No. 157-92 - Intrinsically
Safe and Non-incendive Equipment for Use in Hazardous
Locations

Additional Information:

See the UL Online Certifications Directory at
www.ul.com/database for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's
Certification and Follow-Up Service.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Assistant Chief Engineer, Global Inspection and Field Services

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please
contact a local UL Customer Service Representative at www.ul.com/contactus

