

# CERTIFICATE OF COMPLIANCE

**Certificate Number** 20161221-E62011  
**Report Reference** E62011-20161216  
**Issue Date** 2016-DECEMBER-21

**Issued to:** SKC INC  
863 VALLEYVIEW RD  
EIGHTY FOUR, PA 15330

**This is to certify that  
representative samples of**

**AIR-SAMPLING EQUIPMENT FOR USE IN HAZARDOUS  
LOCATIONS**

Intrinsically safe portable air sampling pump PocketPump  
Touch Model 220-1000TC for use in Class I, Div. 1, Groups  
A, B, C, and D; Class II, Div. 1, Groups E, F, and G; and  
Class III, Div. 1 hazardous locations when used with  
SKC Battery Pack Model P76303.

Have been investigated by UL in accordance with the  
Standard(s) indicated on this Certificate.

**Standard(s) for Safety:**

UL 913, Intrinsically Safe Apparatus and Associated  
Apparatus for Use in Class I, II, and III, Division 1,  
Hazardous (Classified) Locations.


CAN/CSA C22.2 No. 157-92, Intrinsically Safe and Non-  
incendive Equipment for Use in Hazardous Locations.

**Additional Information:**

See the UL Online Certifications Directory at  
[www.ul.com/database](http://www.ul.com/database) for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's  
Certification and Follow-Up Service.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Director North American Certification Program

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please  
contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>

