

CERTIFICATE OF COMPLIANCE

Certificate Number 20190709-E62011
Report Reference E62011-20070830
Issue Date 2019-JULY-09

Issued to: SKC INC
863 VALLEYVIEW RD
EIGHTY FOUR PA 15330

**This certificate confirms that
representative samples of**

AIR-SAMPLING EQUIPMENT FOR USE IN HAZARDOUS
LOCATIONS

See Addendum page

Have been investigated by UL in accordance with the
Standard(s) indicated on this Certificate.

Standard(s) for Safety:

Standard No. UL 913, Intrinsically Safe Apparatus and
Associated Apparatus for Use in Class I, II, and III, Division
1, Hazardous (Classified) Locations
Standard No. CAN/CSA C22.2 No. 157-92, Rev. 2003-06,
(Reaffirmed 2016), Intrinsically Safe and Non-incendive
Equipment for Use in Hazardous Locations


Additional Information:

See the UL Online Certifications Directory at
<https://iq.ulprospector.com> for additional information.

This *Certificate of Compliance* does not provide authorization to apply the UL Mark. Only the UL Follow-Up
Services Procedure provides authorization to apply the UL Mark.

Only those products bearing the UL Mark should be considered as being UL Certified and covered under UL's
Follow-Up Services.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Director North American Certification Program

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please
contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>



CERTIFICATE OF COMPLIANCE


Certificate Number 20190709-E62011
Report Reference E62011-20070830
Issue Date 2019-JULY-09

This is to certify that representative samples of the product as specified on this certificate were tested according to the current UL requirements.

Models

USL Intrinsically safe portable air sampling pump Airchek XR5000, Catalog No. 210-5001, 210-5002, 210-5011, 210-5012, 210-5021, 210-5022, 210-5031 and 210-5032, for use in Class I, Groups A, B, C, and D; Class II, Groups E, F, and G; and Class III hazardous locations when used with SKC Battery Pack, Part Nos. P85002, P85004, P85002A, or P85004A.

CNL Intrinsically safe portable air sampling pump Airchek XR5000, Catalog No. 210-5001, 210-5002, 210-5011, 210-5012, 210-5021, 210-5022, 210-5031 and 210-5032, for use in Class I, Groups A, B, C, and D; Class II, Groups F, and G; and Class III hazardous locations when used with SKC Battery Pack, Part Nos. P85002, P85004, P85002A, or P85004A.



Bruce Mahrenholz, Director North American Certification Program

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>

